

## First retractable lung spreader to facilitate extrapleural dissection of the internal thoracic artery

Yoandy López-de la Cruz<sup>1</sup>✉, MD, MSc; Chenly Pedraza-González<sup>2</sup>, MD; Yolepsis F. Quintero-Fleites<sup>1</sup>, MD, MSc; Roger Mirabal-Rodríguez<sup>1</sup>, MD, MSc; Gustavo de J. Bermúdez-Yera<sup>1</sup>, MD, MSc; Alain Allende-González<sup>1</sup>, MD; Ernesto Chaljub Bravo<sup>1</sup>, MD; and Rodolfo Morales Mato<sup>1</sup>, MD

<sup>1</sup> Department of Cardiovascular Surgery, Hospital Provincial Universitario Cardiocentro Ernesto Che Guevara. Santa Clara, Villa Clara, Cuba.

<sup>2</sup> Faculty of Medicine. Universidad de Ciencias Médicas de Villa Clara. Santa Clara, Villa Clara, Cuba.

*Este artículo también está disponible en español*

### ARTICLE INFORMATION

Received: March 10, 2019

Accepted: April 16, 2019

### Competing interests

The authors declare no competing interests

### ABSTRACT

Internal thoracic artery dissection is one of the most defining procedures of coronary artery bypass grafting, as it has been demonstrated that the quality of life and short-long term outcome of patients will depend on the morphological characteristics of the graft. More than 25 years ago, the use of a malleable lung spreader was proposed to facilitate the internal thoracic artery dissection especially its proximal segment, but that device has never been manufactured. A lung spreader, assembled from disused devices, used in coronary artery revascularization is presented. The main advantages confirmed over more than three years in nearly a hundred patients are discussed and commented upon.

**Keywords:** Myocardial revascularization, Internal thoracic artery, Pleural cavity, Pneumothorax, Chronic obstructive pulmonary disease

### *Primer separador retráctil de pulmón para facilitar la disección extrapleural de la arteria torácica interna*

### RESUMEN

La disección de la arteria torácica interna es uno de los procedimientos más definitivos de la cirugía de revascularización miocárdica, pues se ha demostrado que la calidad de vida y la evolución a corto y largo plazo de los pacientes dependerán de las características morfológicas del injerto. Hace más de 25 años se propuso la utilización de un separador pulmonar maleable para facilitar la disección de la arteria torácica interna, especialmente su segmento proximal, pero nunca se ha fabricado ese dispositivo. Se presenta un separador pulmonar manufacturado a partir de dispositivos en desuso, empleados en cirugía coronaria. Se comentan y discuten las principales ventajas demostradas durante más de tres años en cerca de un centenar de pacientes.

**Palabras clave:** Revascularización miocárdica, Arteria torácica interna, Cavidad pleural, Neumotórax, Enfermedad pulmonar obstructiva crónica

✉ YLópez de la Cruz

Ave 26 de Julio, Edificio 306 Apto. 18.

Reparto Escambray Sur.

Santa Clara, Villa Clara, Cuba.

E-mail address:

yoandyjc@infomed.sld.cu

### INTRODUCTION

At the end of the XIX century, in the Hospital Kasr El Aini, of Cairo, Egypt,

an official doctor of the English imperial army, practically isolated and due to the necessity<sup>1</sup>, conducted for the first time the most widely used of all incisions over the chest cavity. In two articles of the years 1897 and 1901, Herbert Meyrick Nelson Milton (October 1857 – December 20, 1921) described the median longitudinal sternotomy technique to approach to the mediastinum, first for resection of a huge tuberculous lymph node and later for the extraction, through the trachea, of a foreign body lodged in a bronchus<sup>2-4</sup>.

Although, clearly, this incision provides easy access to the organs of the anterior superior mediastinum, it was not until 1953 that Harry B. Shumacker (May 20, 1908 – November 14, 2009) uses it in pulmonary valvotomies closed in eight patients, and he proposed it for cardiac surgery, describing it erroneously as a new approach. However, curiously, the advantages of the exhibition on the median line had not been cheerfully received some time ago, when the American surgeon first submitted his manuscript at the annual meeting of the American Association of Thoracic Surgery<sup>5</sup>, and the preferred incision so far was kept in use during the first decades of surgeries on the heart, which was: the transternal anterior bilateral thoracotomy<sup>6</sup>.

According to Dalton *et al*<sup>7</sup>, in October 1957, Ormand Clinkenbeard Julian (May 6, 1913 – December 18, 1987) restarted the discussion and proposed the reintroduction of the Milton's incision in the incipient heart surgery with extracorporeal circulation, with the fundamental objective of evading pulmonary complications secondary to the opening of both thoracic hemicavities during the "clamshell" approach. He mentioned in his article that the greatest benefit obtained with access to the heart through the breastbone was to avoid any trauma to the lung, especially in the group of cyanotic patients, understood as patients with severe congenital heart disease. The possibility of operating without opening the pleurae would allow the anesthesiologist to manually insufflate the lungs at key moments of the surgery, which avoided alveolar collapse and atelectasis<sup>7</sup>. However, still in 1958, the sternotomy was avoided, and, through incisions in "clamshell", the first anastomosis between mammary and coronary arteries were performed, when Longmire had complications with some of his open endarterectomies<sup>8,9</sup>.


At the end of the decade of 1950<sup>10</sup>, a group from the Cleveland Clinic of Ohio, led by William H. Sewell, proposed the modification of the implantation technique of an internal thoracic artery (mammary)

skeletonized in the myocardium (Vineberg)<sup>11</sup>, by tunneling the vessel, now prepared as a tissue web shaped by the mentioned artery, its veins, endothoracic-pleura parietal fascia, nodes and other perivascular tissues ("pedicle surgery")<sup>12,13</sup>. With the subsequent development of the aorto-coronary graft surgery, the practice of dissecting the pedicle-shaped mammary artery quickly became the technique of choice for most groups, as it is faster and theoretically less harmful to the hemoduct<sup>14-16</sup>.

One of the disadvantages of this procedure is that its first step, usually, consists in the opening of the pleural cavity<sup>17,18</sup>, to properly visualize some segment of the artery that will be taken as the start of the dissection. On a few occasions it is possible to preserve the pleural integrity with this dissection variant, but many surgeons prefer to build a pedicle wide enough to avoid the accidental injury to an artery that does not always remain visible throughout its preparation; in that circumstance, it is practically inevitable not to invade the pleural space.

In patients with a history of nicotine, chronic obstructive pulmonary disease or emphysema, it is practically impossible to access the most proximal portions of the mammary artery without momentarily modifying the anatomy of the surrounding organs, especially hyperinflated lungs that try to occupy their usual position. Some of the solutions implemented in this situation consist in depressing it manually with the separators commonly used in the thoracic surgery, with a long clamp protected with gauze, or –more frequently– with the metal suction cannula; it can also be passively deflated by placing a weight on it, such as wet packs (**Fig. 1**). Another common practice of surgeons is to ask the anesthesiologists to decrease the tidal volume or the positive end-expiratory pressure (PEEP). Any of these strategies can sufficiently upset the anatomy or physiology of the respiratory system, to cause significant complications during and after surgery.

In the early 1970s some groups began to practice, stably, the skeletonized dissection of the mammary artery for the coronary artery bypass grafting (CABG)<sup>19</sup>. This technique allows, frequently, to preserve the integrity of the pleural cavity<sup>20-23</sup>. Nonetheless, also during the approach to the artery segment comprised in the first three intercostal spaces, the ventilation of the lung may seriously complicate the surgeon's handling and involve an injury of the vascular and nerve structures or the pleura and lung, and the use of more time than the desired for dissection.



**Fig. 1. A.** When an emphysematous lung hinders the access to the proximal segment of the mammary, its anatomy must inevitably be modified by depression or separation. **B.** Depression by classic pulmonary spreader. **C.** Depression with the vacuum aspiration metal cannula.

In 1992, Cunningham *et al*<sup>23</sup> indicated that a strategically placed retractable spreader could provide an excellent exposure in this situation. For more than a quarter of a century, this device has not been mentioned again.

In an extensive review of articles related to dissection of the mammary artery in any of its variants, published since the 1940s, only one work was found with the use of hemostats conveniently placed, proposed to optimize extrapleural dissection of the hemoduct, where the authors theorized the future production of more sophisticated spreaders<sup>24</sup>. The analyzed bibliography does not allow the strategies used by other surgical groups to deal with pulmonary interposition during the dissection of the mammary artery.

Since the end of 2015 our group made and began using a retractable lung spreader with the aim of facilitating the skeletal dissection maneuver of the internal mammary artery. To our knowledge, this is the first device manufactured in the world with that intention. Below are its main characteristics and advantages.

### CHARACTERISTICS OF THE PROPOSED SPREADER

The core element of the pulmonary retractable spreader presented here is the articulated arm of the epicardial coronary arteries' stabilizers, used in the CABG. These devices are usually manufactured for a single use, thus, they are discarded very soon. For

this reason, there is a wide availability of this attachment in centers of cardiovascular surgery where coronary surgery is performed without extracorporeal circulation or other beating heart surgical variants.

The "valve" of the spreader is manufactured from aluminum or stainless steel rods; these components, and the small sheets of malleable metal used to join them, are also obtained from instruments discarded in cardiovascular surgery: pleurotomes of small caliber, Overholt or urine-meter container clamps, cardioplegia cannulas, and others. In some of the spreaders we have built, their components have been simply joined with surgical steel suture used for sternal closure (**Fig. 2**). The total manufacturing time, with a minimum of tools, does not take more than one hour.

### ADVANTAGES

The adjustable support design of the coronary stabilizer allows to quickly adhere to a wide variety of sterilizers used during the mammary dissection. It may be convenient to manufacture them in several sizes, so that they can adhere to different anatomical situations, but we have manufactured at least one in which the curved metal rod ("valve") can slide through the plastic supports of the articulated arm, which allows to vary its extension (**Fig. 3**). Once the spreader has been properly fixed to a rigid structure, the surgeon can position it conveniently in a



**Fig. 2.** One of the spreaders manufactured by our team.

way that it does not interfere with its maneuverability (**Fig. 4**). It should be remembered that it is placed almost exclusively when the dissection of the proximal segments of the mammary is hindered by pulmonary expandability, therefore, with only few uses, the surgeon is able to easily predict the most advantageous position that will allow him/her to conclude the procedure without major setbacks.

In situations that need its deployment, the spreader, by retaining some flexibility, reduces or eliminates traumas on the lung, by a gentler depression of the viscera, stably, and only strictly necessary to discover the path of the mammary, with obvious advantages over ventilatory dynamics during that time. This plays an important role in emphysematous patients or with abundant bulbs in the superomedial portions of the lung, which, when insufflated, greatly diminish the surgeon's visualization, and that when separated sharply can be easily injured.

A better visualization of the surgeon obviously facilitates the dissection of the mammary artery in any variant, decreases the duration of the procedure, and the risk of lesion to periarterial structures and the incidence of atelectasis, when re-


ducing trauma and lung without the need to decrease PEEP or pulmonary tidal volume during dissection.

An adjustable device that assumes the separation of anatomical structures will always rationalize the use of surgical team personnel in any intervention; in our case, when eliminating the need for one nurse depressing the lung during preparation of the proximal segment of the mammary, her/his function of instrumentalist is optimized for supporting the extraction of hemoducts that are dissected synchronously (mammary and radial artery, saphenous vein).

On many occasions, a gentle separation of the lung will allow the integrity of the pleural cavity to be preserved, which will result in many benefits, during and –especially– after surgery. In the transoperative period, a closed pleura will prevent the unnoticed accumulation of blood in the



**Fig. 3.** Spreader facilitating dissection of the proximal segment of the left mammary artery. Notice that the curved rod in the form of a “valve” can slide at the surgeon's convenience (arrow), which modifies the scope of the spreader. In this patient, the distal end of the mammary has been pinched at the beginning of the dissection, in line with the new method of mammary dissection proposed by the lead author of this article.



**Fig. 4.** The proper arrangement of the separator does not interfere with the surgeon's maneuverability. In this patient, in addition, the mammary vein has been separated from the artery by a traction point anchored to the edge of the wound, a maneuver that also facilitates the dissection of the proximal segment of the hemoduct.

thoracic cavity, which indirectly forces the surgeon to locate and control inaccurate bleeding points, when the bleeding begins to overflow the anterior mediastinum; this will decrease the possibility of anemia and the need for transfusions. Following this same logic, the intact pleura will prevent the mixing of the papaverine used to irrigate the hemoduct, with amounts of blood that can subsequently be aspirated to the cardiotomy reservoir, and infused back into the patient. This issue has been little studied, but it is theoretically possible that the antispasmodic used during dissection of hemoducts, if incorporated into the bloodstream of the patient, can be responsible for hemodynamically significant transoperative vasoplegia.

The opening of the pleural cavity during mammary dissection may also be responsible for high mor-

bidity during the postoperative period. Pleural effusions have been reported in about 90% of patients in the first week after surgery, some are cataloged as chronic persistent and related to the mammary preparation; and in many cases they require a surgical approach<sup>25</sup>. In addition, the opening of the pleural space can cause pneumothorax and fistulas between systemic arteries and pulmonary vessels, among other complications. The consequent placement of drains through intercostal spaces, contradictorily, can hinder the resolution of effusions by interfering with the absorptive capacity of the pleura<sup>25</sup>, and can be an important source of postoperative thoracic painful syndromes<sup>26,27</sup> causing pneumonia, invasive mechanical ventilation and prolonged stay in the intensive care unit, among other complications.

Finally, it is necessary to mention the wide versatility offered by the device for its use in other specialties where the organs' separation is required without compromising the bimanual skills of surgeons, especially during thoracic surgery, where it can be attached to the classic Finochietto spreader.

## CONCLUSIONS

The evidence shows that the preservation of pleural integrity during the preparation of the internal mammary artery is associated with lower incidence of pulmonary dysfunction and respiratory complications in the postoperative stage<sup>28</sup>; fully commenting on the benefits of this strategy is beyond the objectives of this work.

In cardiovascular surgery, patients in which the integrity of the pleural cavity can be preserved will dispense with the classic intercostal drainage, and even occasionally, a single probe may be placed in the anterior mediastinum<sup>16</sup>. The chances of developing unilateral pleural effusion, infectious respiratory complications, intercostal neuralgia or atelectasis will be minimal, as well as bleeding and the need for transfusions. An intact pleura will significantly reduce the formation of pulmonary adhesions and the distortion of the intrathoracic anatomy<sup>29</sup>, which is extremely beneficial for the patient and surgeon if the need for an oncological surgical procedure subsequently arises, an increasingly frequent situation in the current thoracic surgery<sup>30,31</sup>.

To all this, we must add the well-known benefits

of the skeletonized dissection of the mammary artery in virtually any revascularization strategy that it is employed, relative to the low incidence of thoracic pain or paresthesia<sup>32,33</sup>, and deep complications of sternal wound<sup>34-36</sup>.

Since the end of 2015, the lung spreader has been used in about one hundred patients operated in our department. A preliminary analysis of its effectiveness showed transoperative benefits in relation to the conservation of pleural integrity, decreased dissection time and protection of the mammary vein<sup>28</sup>, the latter situation that can also be translated into greater protection of the sternum, because some authors have pointed out the existence of a possible theoretical relationship between retrosternal venous drainage injury and deep bone complications<sup>36,37</sup>. Perhaps it should also be noted that during all this time there have been no cases of mediastinitis in our department, a complication with known medical and economic repercussions<sup>38,39</sup>. In addition, as overall standard, the patients in which the device has been used have been fewer days in the postoperative intensive care unit and have been discharged more precociously, which has undoubtedly meant a substantial reduction in hospital costs in this important group of patients.

## REFERENCES

1. van Wingerden JJ. Sternotomy and intrathoracic omentum: two procedures, two innovators, and the river that runs through it - A brief history. *Ann Thorac Surg*. 2015;99(2):738-43.
2. Milton H. Mediastinal surgery. *Lancet*. 1897; 149(3839):872-5.
3. Milton H. Removal of a foreign body from the bronchus by intrathoracic tracheotomy. *Lancet*. 1901;157(4039):242-4.
4. Shiffman MA, Low M. *Recent Clinical Techniques, Results, and Research in Wounds*. Berlín: Springer International Publishing; 2018.
5. van Wingerden JJ. *Poststernotomy mediastinitis revisited*. Amsterdam: Optima Grafische Communicatie; 2015.
6. Bains MS, Ginsberg RJ, Jones WG, McCormack PM, Rusch VW, Burt ME, *et al*. The Clamshell incision: An improved approach to bilateral pulmonary and mediastinal tumor. *Ann Thorac Surg*. 1994;58(1):30-3.
7. Dalton ML, Connally SR, Sealy WC. Julian's reintroduction of Milton's operation. *Ann Thorac Surg*. 1992;53(3):532-3.
8. Longmire WP, Cannon JA, Kattus AA. Direct-vision coronary endarterectomy for angina pectoris. *N Engl J Med*. 1958;259(21):993-9.
9. López de la Cruz Y. William Polk Longmire Jr. y los primeros 60 años de la cirugía de puentes aorto-coronarios. *CorSalud* [Internet]. 2018 [citado 2 Mar 2019];10(2):158-63. Available at: <http://www.revcorsalud.sld.cu/index.php/cors/article/view/338/691>
10. Sewell WH. The surgical treatment of coronary artery disease. *Am Surg*. 1959;25:560-6.
11. Kieser TM, Rose MS, Aluthman U, Narine K. Quicker yet safe: skeletonization of 1640 internal mammary arteries with harmonic technology in 965 patients. *Eur J Cardiothorac Surg* [Internet]. 2014 [citado 4 Mar 2019];45(5):e142-50. Available at: <https://academic.oup.com/ejcts/article-pdf/45/5/e142/1135956/ezu024.pdf>
12. Sewell WH, Sones FM, Fish RG, Joyner JT, Effler DB. The pedicle operation for coronary insufficiency: Technique and preliminary results. *J Thorac Cardiovasc Surg*. 1965;49:317-29.
13. Mueller RL, Rosengart TK, Isom OW. The history of surgery for ischemic heart disease. *Ann Thorac Surg*. 1997;63(3):869-78.
14. Benedetto U, Altman DG, Gerry S, Gray A, Lees B, Pawlaczyk R, *et al*. Pedicled and skeletonized single and bilateral internal thoracic artery grafts and the incidence of sternal wound complications: Insights from the Arterial Revascularization Trial. *J Thorac Cardiovasc Surg*. 2016;152(1):270-6.
15. Satdhabudha O, Noppawinyoowong N. A randomized comparison of flow characteristics of semiskeletonized and pedicled internal thoracic artery preparations in coronary artery bypass. *J Cardiothorac Surg* [Internet]. 2017 [citado 5 Mar 2019];12(1):28. Available at: <http://doi.org/10.1186/s13019-017-0589-1>
16. Özülkü M, Aygün F. Effect of LIMA harvesting technique on postoperative drainage in off-pump CABG. *Braz J Cardiovasc Surg*. 2016;31(2):120-6.
17. Badia S, Berastegui E, Cámara ML, Delgado L, Fernández C, Julià I, *et al*. Revascularización miocárdica con uso de doble arteria mamaria interna y morbilidad esternal. Experiencia de un centro. *Cir Cardiovasc*. 2017;24(3):149-56.
18. Raffa GM, Malvindi PG, Kowalewski M, Sansone

- F, Menicanti L. Training in coronary artery bypass surgery: Tips and tricks of the trade. *Semin Thorac Cardiovasc Surg.* 2017;29(2):137-42.
19. Galbut DL, Traad EA, Dorman MJ, DeWitt PL, Larsen PB, Kurlansky PA, *et al.* Seventeen-year experience with bilateral internal mammary artery grafts. *Ann Thorac Surg.* 1990;49(2):195-201.
  20. Uzun A, Yener AÜ, Kocabeyoglu S, Çiçek ÖF, Yasar E, Yener Ö, *et al.* Effects of pleural opening on respiratory function tests in cardiac surgery: A prospective study *Eur Rev Med Pharmacol Sci.* 2013;17(17):2310-7.
  21. Bortolussi G, Bejko J, Gallo M, Comisso M, Carrozzini M, Guglielmi C, *et al.* Coronary artery bypass grafting in elderly patients: Insights from a comparative analysis of total arterial and conventional revascularization. *J Cardiovasc Transl Res.* 2016;9(3):223-9.
  22. Di Mauro M, Iacò AL, Acitelli A, D'Ambrosio G, Filippini L, Salustri E, *et al.* Bilateral internal mammary artery for multi-territory myocardial revascularization: long-term follow-up of pedicled versus skeletonized conduits. *Eur J Cardiothorac Surg.* 2015;47(4):698-702.
  23. Cunningham JM, Gharavi MA, Fardin R, Meek RA. Considerations in the skeletonization technique of internal thoracic artery dissection. *Ann Thorac Surg.* 1992;54(5):947-50.
  24. Efthymiou CA, Weir WI. Optimisation of internal thoracic artery exposure using a simple retraction method for extrapleural dissection. *Ann R Coll Surg Engl.* 2015;97(2):158.
  25. Charniot JC, Zerhouni K, Kambouchner M, Martinod E, Vignat N, Azorin J, *et al.* Persistent symptomatic pleural effusion following coronary bypass surgery: clinical and histologic features, and treatment. *Heart Vessels.* 2007;22(1):16-20.
  26. Dabbagh A. Postoperative Pain Management in Cardiac Surgery. En: Dabbagh A, Esmailian F, Aranki SF, eds. *Postoperative Critical Care for Cardiac Surgical Patients.* Heidelberg: Springer; 2014. p. 257-94.
  27. Schoeffler P, Chabanne R. Surgical Factors Influencing the Occurrence of Chronic Postsurgical Pain. En: Mick G, Guastella V, eds. *Chronic Postsurgical Pain.* Heidelberg: Springer; 2014. p. 59-71.
  28. López-de la Cruz Y, Pedraza-González C, Quintero-Fleites YF, Mirabal-Rodríguez R, Bermúdez-Yera GJ, Allende-González A, *et al.* Beneficios inmediatos de la utilización de un separador pulmonar flexible durante la disección esquelizada de la arteria mamaria interna. *Cir Cardiovasc.* 2019; 26(2):98-103.
  29. Oates M, Yadav s, Saxena P. Management of a left internal thoracic artery graft injury during left thoracotomy for thoracic surgery. *Heart Lung Circ [Internet].* 2016 [citado 7 Mar 2019];25(7):e75-7. Available at: <http://dx.doi.org/10.1016/j.hlc.2016.01.008>
  30. Kanzaki R, Inoue M, Minami M, Shintani Y, Funaki S, Kawamura T, *et al.* Outcomes of lung cancer surgery in patients with coronary artery disease: a decade of experience at a single institution. *Surg Today.* 2017;47(1):27-34.
  31. Maeda H, Kanzaki M, Sakamoto K, Isaka T, Yamazaki K, Onuki T. Surgery for non-small cell lung cancer in patients with a history of cardiovascular surgery. *Surg Today.* 2017;47(3):284-92.
  32. Bawany FI, Khan MS, Khan A, Hussain M. Skeletonization technique in coronary artery bypass graft surgery reduces the postoperative pain intensity and disability index. *J Card Surg.* 2014; 29(1):47-50.
  33. Saggio G, Bianco A, Orenzo G, Tancredi G, Del Gaudio C, Zeitani J. An Electronic-Engineered Sensory Sternal Retractor Aimed at Post-sternotomy Pain Reduction. En: Peixoto N, Silveira M, Ali H, Maciel C, van den Broek E, eds. *Biomedical Engineering Systems and Technologies. BIOSTEC 2017.* Cham: Springer; 2018. p. 3-18.
  34. Sajja LR. Strategies to reduce deep sternal wound infection after bilateral internal mammary artery grafting. *Int J Surg.* 2015;16(Pt B):171-8.
  35. Ohira S, Miyata H, Yamazaki S, Numata S, Motomura N, Takamoto S, *et al.* Deep sternal wound infection after bilateral internal thoracic artery grafting: Insights from a Japanese national database. *J Thorac Cardiovasc Surg.* 2019;157(1):166-73.
  36. Ding WJ, Ji Q, Shi YQ, Ma RH, Wang CS. Incidence of deep sternal wound infection in diabetic patients undergoing off-pump skeletonized internal thoracic artery grafting. *Cardiology.* 2016; 133(2):111-8.
  37. Wehman B, Taylor B. Coronary revascularization using bilateral internal thoracic arteries: Safe with skeletonization? *J Clin Exp Cardiol [Internet].* 2013 [citado 8 Mar 2019];(Supl 7):007. Available at: <http://doi.org/10.4172/2155-9880.S7-007>
  38. Fu RH, Weinstein AL, Chang MM, Argenziano M, Ascherman JA, Rohde CH. Risk factors of infected sternal wounds versus sterile wound dehiscence.

J Surg Res. 2016;200(1):400-7.

39. Narayan P, Sarkar K, Trehan N, Chandra P, Chouhan NS, Puskas JD, *et al*. Key updates from inter-

national coronary congress 2016 – A review. Indian J Thorac Cardiovasc Surg. 2018;34(1):44-52.