



Evaluation of data balancing techniques. Application to CAD of lung nodules using the LUNA16 framework

Alexander Sónora-Mengana^(*), Evgenia Papavasileiou^(*), Juan C. García-Naranjo, Bart Jansen, Jef Vandemeulebroucke

RESUMEN/ABSTRACT

Due to the high incidence of lung cancer, computer-aided detection (CAD) systems may play an increasingly important role in screening. Classification in CAD systems has to deal with highly imbalanced datasets composed by actual nodules and non-nodule structures. The application of data balancing techniques helps the training process of the classifiers, making the generation of the classification rules more effective. The purpose of this paper is to compare the performance of different data balancing techniques applied to the classification of lung nodules. According to the reviewed literature, this is the first time that different data balancing methods are evaluated on the problem of lung nodule detection using a large data set and at low false positive rates. A web-based framework was used to evaluate the different methods applied to a classical CAD system (ETROCAD) presented in the LUNA16 Challenge by calculating a score of average sensitivity at different values of false positives per scan. In our experiments, data balancing using SMOTE and SMOTE-TL led to the best results, with a score of 0.760 and 0.759 respectively, in comparison to 0.748 when not balancing the data. Although the impact on the overall score may seem marginal, adequate data balancing resulted in the correct classification of 36 additional candidate nodules at 4 FP/scan. At the time of writing this paper, the SMOTE-based ETROCAD system had the best score among all the classical systems using handcrafted features in LUNA16 web site.

Keywords: Data balance, Computer Aided Detection, Near-Miss, CNN, Random Under-sample, Tomek links, Self-Organized Map, Random Over-sample, ADASYN, SMOTE, LUNA16

Debido a la alta incidencia del cáncer de pulmón a nivel mundial, los sistemas de diagnóstico asistidos por computadora (CAD por sus siglas en inglés) desempeñan un papel importante en los estudios de pesquiasaje de la enfermedad. El proceso de clasificación en los sistemas CAD se ve deteriorado debido al bajo porcentaje de estructuras detectadas que se corresponden a nódulos verdaderos. El principal propósito de este trabajo es compararla influencia de las técnicas de balanceo de datos en la clasificación de nódulos pulmonares. De acuerdo con la literatura revisada, en este trabajo se presenta por primera vez la comparación entre balanceo de datos aplicado a la detección de nódulos pulmonares empleando un conjunto de imágenes grande para razones de falsos positivos bajas. Los métodos se aplicaron a un sistema CAD presentado en LUNA16 Challenge (ETROCAD). Los mejores resultados obtenidos se corresponden a los métodos SMOTE y SMOTE-TL con una sensibilidad promedio de 0.760 y 0.759 respectivamente, en contraste a 0.748 obtenido sin realizar ningún balanceo de datos. Aunque el impacto en el índice empleado en LUNA16 no es alto, un balanceo de datos adecuado permitió la detección correcta de 36 candidatos adicionales a una raso de 4 falsos positivos por imagen. En el momento de escritura de este trabajo, el desempeño del ETROCAD con balanceo de datos basado en SMOTE exhibe la mayor puntuación de entre los sistemas CAD clásicos.

Palabras clave: balanceo de datos, sistemas de detección asistidos por computadora, Near-Miss, CNN, submuestreo aleatorio, Tomek links, Mapas auto-organizados, sobremuestreo aleatorio, ADASYN, LUNA16

Evaluación de técnicas de balanceo de datos. Aplicación en un sistema CAD para nódulos pulmonares empleando el sistema LUNA16

*These authors contributed equally to this work

1.-INTRODUCTION

Many real-life classification problems lead to situations in which the distribution of the positive and negative examples is unbalanced, because individuals of some classes appear more frequently. This is known as class imbalance and it is not uncommon to have an imbalance of several orders of magnitude. As an example, Tan et al. [1] proposed a CAD system for lung nodule classification, in which they classified 111 332 candidate structures, of which only 574 structures are actual nodules. From a machine learning point of view, this imbalanced nature of the training data represents a challenge for the learning algorithm, as it biases the outcome towards the majority class, i.e. the FPs [2].

In machine learning, several methods have been proposed to handle the class imbalance [3]. A discussion of open issues and present challenges for further development in the field of imbalanced learning can be found in [4]. Typically, to provide a balanced distribution, sampling methods are used. The sampling methods can be classified in under-sampling methods or over-sampling methods, depending on if they remove or add samples to one of the classes, respectively. The combination of under-sampling and over-sampling is referred as ensemble methods.

For a clear presentation in the rest of this work, some definitions are presented. Consider a training set $S = S_{maj} \cup S_{min}$, where S_{maj} and S_{min} are the majority and minority class respectively, such that $S_{maj} \cap S_{min} = \{\emptyset\}$. The samples $x_i \in S_{maj}$ and $y_j \in S_{min}$ will denote individuals of the majority and minority classes, respectively. Finally, any set generated from sampling methods on S are labeled as R .

1.1.-SAMPLING METHODS FOR DATA IMBALANCE

A first group of methods resample the classes in a random fashion, either by augmenting the S_{min} or reducing the size of S_{maj} . The random over-sampling method augments S_{min} replicating randomly selected examples. In this way, the class distribution balance is adjusted. On the other hand, random under-sampling works by removing randomly selected samples from S_{maj} instead of adding samples. These two mechanisms provide a way for varying the degree of class balance in any desired way. Despite of its simplicity, each method is associated with issues that can compromise the performance of the learning algorithm [5], [6], [7].

The drawback of random under-sampling is obvious as the removal of samples from S_{maj} can cause the classifier to miss useful information. Inversely, the over-sampling method can lead to overfitting, due to the replication of some of the samples from S_{min} . Overfitting could make the decision rules too specific, with high accuracies on the training set, but poor performance on the testing set [5].

Condensed nearest neighbor (CNN) samples S_{maj} , eliminating random samples without significantly affecting the performance of the nearest neighbor classification, i.e. the nearest neighbor rule used to classify S_{maj} should give almost the same result if applied to the under-sampled set R [8]. This is especially true if the elements in R are representative elements of the S_{maj} .

Tomek links (TL) is a data cleaning technique used to remove the overlap introduced by sampling methods. A Tomek link is a pair (x_i, y_j) such that x_i and y_j are minimally distanced nearest neighbors. Well-defined class clusters (R) can be established removing all x_i from the Tomek links until all minimally distanced nearest neighbor pairs belong to the same class. This method will lead to well-defined classification rules for improved classification performance [3].

The Near Miss methods are other data cleaning techniques for under-sampled data sets. Three variants of Near-Miss are presented in [9]. NearMiss-1 (NM1) selects the x_i neighbors closest to some determined number of y_j and removes them if the average distance is minimal. NearMiss-2 (NM2) works in a similar manner but considering all the y_j , and taking the average distance to a determined number of the farthest y_j . The third method, NearMiss-3 (NM3), selects a given number of x_i , surrounding each y_j . The performance of these methods could be influenced by the distribution of the x_i among the y_j [9].

The Synthetic Minority Oversampling Technique (SMOTE) generates synthetic examples operating in feature space [10]. S_{min} is over-sampled by taking each y_j and introducing synthetic examples along the line segments joining all of their nearest neighbors. Depending on the amount of required over-sampling, samples from the k-nearest neighbors are randomly chosen. Despite its great success in several applications, it has some drawbacks including over-generalization.

The Adaptive Synthetic (ADASYN) approach adaptively creates different amounts of synthetic data according to their distributions [11]. It uses a density distribution as a criterion to automatically decide the number of synthetic samples that need to be generated for each y_j by adaptively changing the weights of different x_i to compensate for the skewed distributions.

A Self-Organizing Map (SOM) aims to reduce the class imbalance by first clustering all x_i samples into different cells of a Kohonen layer based on the similarities in the feature space. The subset R is generated by taking a number of examples from

each cell. In this way, the size of S_{maj} is reduced without altering the distribution in the feature space. This method was employed by Tan et al. [1] to reduce the class imbalance of the lung nodule classification problem to a factor three.

There is a third category, referred as ensemble methods, that combines over-sampling and under-sampling. Two examples of ensemble methods are SMOTE-TL and SOM-SMOTE. In SMOTE-TL, S_{min} is oversampled with SMOTE to the size of S_{maj} , then both classes are down-sampled to the original size of S_{min} using TL. On the other hand, SOM-SMOTE, first down-sample S_{maj} to a size greater than the size of S_{min} , and then S_{min} is oversampled to the new S_{maj} size. As SOM-SMOTE reduces the sizes the classes as the first step, balanced data sets with less computational load can be obtained.

1.2.-CAD OF LUNG CANCER

Lung cancer has been the most common type of cancer for several decades and the second cause of death worldwide among all non-communicable diseases¹. Also, its detection in initial states is considered the most effective way to improve survival of patients, in which case, the 5-year survival rate is approximately 54 %, compared to 4 % in case of detection in advanced stages [12]. These facts have led to the development of screening programs in United States of America [13,14] and Europe². The implementation of the programs at large scale will increase the radiologist burden significantly and the computer-aided diagnosis (CAD) systems may play a significant role in reducing the reading time and thereby improving cost-effectiveness [15,16].

A typical CAD approach for lung cancer detection has three steps: preprocessing, candidate detection, and false positive reduction. The preprocessing is used for image standardization (spatially resampling the image to isotropic and homogeneous resolution) and lung segmentation in order to restrict the search space. The candidate detection step aims to detect nodule candidates at a very high sensitivity. As a consequence, a large imbalance between true positives (TP) and false positives (FP) is generated. Subsequently, the FP reduction step lowers the number of FPs among the candidates and generates the final set of detected nodules. Figure 1 depicts the CAD system presented in [1], where the three steps of the procedure are shown.




Figure 1

Block diagram for the CAD system presented by Tan et al in [1].

1.3.-THE LUNA16 CHALLENGE

¹ Global Health Estimates 2015: Burden of disease by Cause, Age, Sex, by Country and by Region, 2000-2015. Geneva, World Health Organization; 2016. http://www.who.int/healthinfo/global_burden_disease/en/

²[http://www.thelancet.com/journals/lanonc/article/PIIS1470-2045\(17\)30861-6/fulltext](http://www.thelancet.com/journals/lanonc/article/PIIS1470-2045(17)30861-6/fulltext)

To allow the comparison of different CAD systems of lung nodules, an online framework for the evaluation was introduced in [17]. The LUNg Nodule Analysis 2016 (LUNA16) Challenge has a web interface to evaluate algorithms and compare the results with other approaches. LUNA16 consists of two separate tracks: the full pipeline of a CAD system (track 1) for thoracic CT images, including the candidate nodule detection and the FP reduction; and a simplified track (track 2) covering only the FP reduction for which coordinates of candidate nodules are given. In both tracks, the data originates from the LIDC-IDRI database, consisting of a subset of 888 scans for which reference annotations from four radiologists are available [18]. This data set can be used for training and system evaluation.

In order to evaluate the systems, the Free-response Receiver Operating Characteristic (FROC) analysis is computed. TP are defined as nodules annotated by the majority of the radiologists (at least 3 out of 4 radiologists). Sensitivity is defined as the fraction of detected TP divided by the number of nodules in the LUNA16's reference standard. In the FROC curve, sensitivity is plotted as a function of the average number of FP per scan (FPs/scan). Evaluation is performed using predefined subsets for 10-fold cross-validation, which can be downloaded from the LUNA16 website. As overall score, the average of the sensitivity at 0.125, 0.25, 0.5, 1, 2, 4, and 8 FPs per scan, over all folds, is used. Also, the 95% confidence interval using bootstrapping with 1000 bootstraps is calculated. For each bootstrap, a new set of candidates are constructed using (scan-level) sampling with replacement. A Python evaluation script is available for download on the LUNA16 website and the results can be viewed by all participants.

1.4.-CAD SYSTEMS AND DATA BALANCE

Classifiers with balanced data sets tend to perform better, in particular in terms of increasing sensitivity at low FP rates [10]. As mentioned, CAD systems typically face this problem as well. Despite this, few data balance studies for CAD of lung nodules have been reported. Table 1 shows an overview of previous work in classical CAD system development.

Table 1

Data sets and data balance methods reported in classical CAD systems for lung nodule detection.

	Number of scans	Data balance method		Number of scans	Data balance method
Tan et al. [1]	450	SOM	Han et al. [19]	1012	-
Abbas [20]	220	-	Madgy et al. [21]	80	-
Nithila and Kumar. [22]	106	-	Orozco et al. [23]	45	-
Cirujeda et al. [24]	95	-	Badura and Pietka [25]	23	-
Jacobs et al. [26]	888	-	Sun et al [27]	360	SMOTE
Firmino et al. [28]	420	-	Keshani et al. [29]	50	-
Demir and Çamurcu. [30]	95	-	Shiju et al. [31]	108	SMOTE
Gong et al. [32]	189	SMOTE	Jingchen et al. [33]	1010	SMOTE
Demir and Çamurcu. [34]	200	-	Makaju et al. [35]	1018	-
Baboo and Iyyapparaj [36]	1028	-	Cao et al. [37]	1012	MKFSOS
Sui et al. [38]	120	RU-SMOTE	Mehre et al. [39]	97	SMOTE

As can be seen, most of the reviewed works on classical CAD systems for lung nodules do not report the use of any data balancing methods. Few authors mention data balancing, but did not examine the impact with respect to unbalanced data. To the best of the authors knowledge, only Sui et al. [38] described the impact of data balancing for CAD of lung nodules. In their experiments, 500 candidate nodules from 120 patients, with a class imbalance of 1/6, were classified based on eight features using several techniques to deal with class imbalance. They found that considerable improvement can be obtained by alleviating the data imbalance -the best results obtained using a combination of under- and oversampling together with a biased SVM approach.

The purpose of this paper is to compare the performance of different data balancing techniques applied to the classification of a highly imbalanced biomedical dataset and their impact on nodule classification tasks. With respect to previous work, the presented experiments are based on a larger data set, have higher feature dimensionality, include more data balancing techniques and evaluate sensitivities at both low and high FP rates.

2.-MATERIAL AND METHODS

2.1.-THE ETROCAD SYSTEM

A modification of the system proposed in [1], referred in the rest of this paper as ETROCAD is used to evaluate the effect of data balancing techniques on the performance of a CAD system. The ETROCAD system is modified in terms of using a Support Vector Machine (SVM) with radial basis function kernel. SVMs are well-established classifiers, proven to be useful in many cancer classification tasks [1, 28-40].

The 888 scans from the LUNA16 framework were processed by the first two steps of the approach in Figure 1 (preprocessing and candidate selection) to extract all the candidates. A feature vector is built by features calculated for each candidate. Table 2 shows a short summary of the used features [1].

Table 2
Description of features used in ETROCAD

Feature	Notes
Volume	Equivalent to the number of voxels
Min_diam = $\min(\text{dim}_i)$	Dim _i =minimum diameter corresponding to the principal axis i of the minimum volume-enclosing ellipsoid
Max_diam = $\max(\text{dim}_i)$	
Compactness1	Ratio between volume and product of all dim _i
Compactness2	Ratio between volume and dim_i^3
Elongation factor	Max_dim/min_dim
Bounding ellipsoid feature	1 for 3D ellipsoid, 0 for 2D ellipsoid
Distance to lung wall	From nodule candidate centroid to lung wall
Distance to the center of the slice	From candidate centroid to the center of the slice on which centroid is located
Average of L _{uu} and L _{vv}	On segmented voxels, and spherical kernels of radius 1 and 3 pixels at scales 1 and 2. L _{uu} and L _{vv} features over a gauge coordinate system [1].
Average of nodule filter values	On segmented voxels, and spherical kernels of radius 1 and 3
Average of vessel filter values	On segmented voxels, and spherical kernels of radius 1 and 3
Average of divergence values	On segmented voxels, and spherical kernels of radius 1 and 3
Grey-value features	Mean, median, maximum, minimum and standard deviation on segmented voxels, and spherical kernels of radius 1 and 3

2.2.-BALANCING THE DATA

The SOM-based method was implemented using a SOM Toolbox for Matlab³. In this case, the TP/FP ratio was 1/3 to be consistent with the original system [1]. The rest of data balancing methods and classification were implemented in Python 2.7, using the Imbalanced-learn [41] and the Scikit-learn⁴ modules respectively. Scikit-learn provides easy-to-use tools for data mining and data analysis. Table 3 summarizes the methods implemented in Scikit-learn that were used on the lung nodule data provided by the ETROCAD system. In all the cases, the goal was to get a well-balanced dataset with a final TP/FP ratio approximating 1. The remaining parameters of the methods were set to recommended default values.

2.3.-DATA SET PROCESSING AND EVALUATION

To evaluate the different methods, the feature vectors provided by ETROCAD were balanced by each of the data balancing methods from Table 3. The result was then used to train the ETROCAD's classifier.

The performance of each method was evaluated using the methodology provided by the LUNA16 framework (track 1). The 10-fold cross-validation was run as described in LUNA16 procedure, and then the FROC curve and the score were calculated (average of sensitivities at 0.125, 0.25, 0.5, 1, 2, 4, and 8 FPs per scan). To be able to compare the methods, the evaluation

³<http://www.cis.hut.fi/projects/somtoolbox>

⁴<http://scikit-learn.org/stable/index.html>

script from LUNA16 was modified to extract all the FROC curves from bootstrapping; and all the scores for each one were calculated.

Table 3
Sampling methods and default values for methods used in imbalance learning.

Method	Parameter name	Default value
RUS	No parameters	-
TL	No parameters	-
NM-3	Number of neighbors to be considered	3
CNN	No parameters	-
ROS	No parameters	-
SOM	FP-TP ratio	3
SMOTE	Number of nearest neighbors to construct the synthetic samples	5
ADASYN	Number of nearest neighbors to construct the synthetic samples	5
SMOTE-TL	Number of nearest neighbors to construct the synthetic samples	5
SOM-SMOTE	FP-TP ratio	3
	Number of nearest neighbors to construct the synthetic samples	5

Kruskal-Wallis test is used to find if there are differences at the 1 % of significance level. If differences are found, a multiple comparison is done using the Bonferroni method. Kruskal-Wallis test and multiple comparisons are done using *kruskalwallis* and *multcompare* functions respectively from Matlab’s Statistical and Machine Learning Toolbox.

3.-RESULTS

At the output of the candidate detection phase, ETROCAD detected 435 303 structures (candidates) in 888 scans, of which 1186 correspond to actual nodules according to the ground truth annotations provided. As such, the initial imbalance ratio was approximately 367. Figure 2 shows the graphical output obtained with LUNA16 script for SMOTE. The solid line represents the FROC curve for the system with the balance method and dashed lines mark the limits for a 95 % confidence interval of the bootstrapping procedure.




Figure 2

LUNA16 script output for SMOTE. Solid line: FROC curve for the 10-fold cross-validation, dashed lines: limits for the 95 % confidence interval from bootstrapping. Mean curve is overlapped with the solid line.

From the modified LUNA16 script, 1000 FROC curves were extracted from the bootstrapping procedure for each balancing method. Table 4 shows the mean scores and standard deviations for all sets of curves. Differences at 1 % of significance level were found among the populations using the Kruskal-Wallis test (p-value = 0).

Table 4

Mean scores and standard deviations for all curves from bootstrapping

	No Balance	SOM	CNN	RUS	NM-3	TL	ROS	ADASYN	SMOTE	SMOTE-TL	SOM-SMOTE
Mean score	0.748	0.704	0.55	0.655	0.422	0.749	0.743	0.730	0.760	0.760	0.745
Standard deviation	0.018	0.023	0.018	0.019	0.017	0.018	0.020	0.019	0.019	0.018	0.019

Figure 3 shows the graphical output from the Matlab’s *multcompare* function. This graph shows the estimates and a pair wise comparison in which two groups are significantly different if their intervals (horizontal lines) are disjoint. In this way, only SOM-SMOTE and TL, have no significant differences with NoBalance. The best overall performance was obtained using SMOTE and SMOTE-TL methods.




Figure 3

Differences among methods using multiple comparison test. Because Kruskal-Wallis test was used, horizontal axis corresponds to the ranks of the median for each method. SOM-SMOTE and TL (gray lines) have no significant differences respect to NoBalance (blue line).

A more detailed analysis of the sensitivities at different FP rates is shown in Table 5, along with final average score for each method. For the lowest FP rate (0.125FP/scan), all evaluated methods except for TL have scores below the system with no data balancing. All over-sampling and ensemble methods achieve higher overall scores than the under-sample ones.

Table 5

Sensitivities at selected FP per scan and scores as defined in LUNA16 Challenge

FP/Scan	No Balance	SOM	RUS	CNN	NM-3	TL	ROS	ADASYN	SMOTE	SMOTE-TL	SOM-SMOTE
0.125	0.578	0.245	0.344	0.384	0.294	0.581	0.486	0.453	0.540	0.538	0.515
0.250	0.663	0.509	0.488	0.431	0.332	0.663	0.618	0.590	0.650	0.652	0.628
0.500	0.718	0.647	0.606	0.485	0.368	0.718	0.724	0.707	0.740	0.738	0.728
1.000	0.773	0.750	0.696	0.535	0.420	0.775	0.783	0.780	0.792	0.793	0.778
2.000	0.805	0.810	0.772	0.594	0.458	0.810	0.834	0.824	0.839	0.838	0.824
4.000	0.837	0.854	0.822	0.672	0.513	0.836	0.864	0.862	0.868	0.864	0.862
8.000	0.860	0.886	0.858	0.755	0.581	0.860	0.885	0.882	0.890	0.889	0.891
Score	0.748	0.672	0.655	0.551	0.424	0.749	0.742	0.728	0.760	0.759	0.747

Figure 4 shows the FROC curves for all methods. As can be seen, curves for NM3, CNN and RUS are always below NoBalance. SOM goes over NoBalance only at FP rates higher than 2FP/Scan. The sensitivity values for TL are almost the same than NoBalance. On the other hand, the rest of the methods but ADASYN go higher than NoBalance for FP rates higher than 0.5 FP/Scan.




Figure 4

FROC curves for each method. SMOTE and SMOTE-TL as well as TL and NoBalance are overlapped.

To highlight the impact of data balancing, Table 6 shows a summary of the results obtained at 4 FP/scan, a value commonly used in a clinical setting. For the best method, data balancing resulted in 36 additional nodules correctly classified.

Table 6

Summary for the selected ETROCAD versions. The first row corresponds to the system with no balance method; the other rows are sorted in ascendant order according to sensitivity (Sens). dTP, dTN, dFP and dFN are the detection differences respect to the no balanced system. Negative numbers represent less detections than NoBalance. All values are reported at 4 FP/scan

Method	TP	TN	FP	FN	dTP	dTN	dFP	dFN	Sens (%)	Spec (%)	Acc (%)
NoBalance	992	80025	3434	194	-	-	-	-	83.73	5.35	1.40
NM3	608	80774	3547	578	-384	749	113	384	51.35	14.01	1.39
CNN	796	80227	3538	390	-196	202	104	196	67.20	9.93	1.40
RUS	974	79685	3522	212	-18	-340	88	18	82.21	5.68	1.41
TL	991	79882	3545	195	-1	-143	111	1	83.64	5.21	1.40
SOMSMOTE	1022	74648	3550	164	30	-5377	116	-30	86.17	4.42	1.49
SOM	1023	78623	3551	163	31	-1402	117	-31	86.26	4.39	1.42
ADASYN	1023	79479	3487	163	31	-546	53	-31	86.34	4.47	1.41
SMOTETL	1026	79505	3500	160	34	-520	66	-34	86.59	4.37	1.41
ROS	1028	79495	3520	158	36	-530	86	-36	86.76	4.30	1.41
SMOTE	1028	79492	3533	158	36	-533	99	-36	86.76	4.28	1.41

4.-DISCUSSION

Under sampling methods were found to perform badly in the experiments. Both, RUS and NM3 remove samples without considering whether they are “good examples” of FP. This may have hampered the performance of the system, which is confirmed by the reduced score of both methods. On the other hand, as TL tries to get a well-defined set for training by removing boundary or noisy samples, it is expected to perform better than other methods. For the cases of SMOTE and SMOTE-TL, they were found significant better than the rest of the methods, as can be seen from the Table 4 and Figure 3, although the relative difference in the score is only 0.1 % with respect to no balancing.

When considering the LUNA16 score, differences between the methods are relatively small in terms of average sensitivity, (Table 6). This can be attributed to averaging of sensitivities at different FP rates, and the high number of candidate structures. At 4 FP/scan, and when considering the absolute number of detections, the difference is more apparent.

The performance of the employed CAD system is comparable to the state of the art. Most studies report sensitivities at high FP rates. Achieving high sensitivity at low FP rates is technically challenging, but more suitable for clinical use, which is why it was included in current study. Mehre et al. [39] report sensitivities of 92.91 % at 3 FP/scan. Although this value is higher than the values shown at Table 4, their performance at lower FP rates is worse. To be able to get the values at lower FP rates, the sensitivities at 0.125, 0.25, 0.5, 1, 2, 4, 8 FP/scans from the FROC curve in [39] were estimated using WebPlotDigitizer⁵. The estimated sensitivities were 0.2, 0.34, 0.53, 0.73, 0.87, 0.96 and 0.98; yielding a LUNA16 score of 0.66, which is lower than the best scores reported in Table 5.

The improvement for lung nodule classification when using data balancing reported by Sui et al. [38], was larger with respect to the results obtained in this study. Detailed comparison is however difficult, as differences could be due to the different dataset, the limited amount of scans used in latter study (120 scans) and the low-dimensional feature space.

The impact of proper data balancing can be seen when considering the results of the LUNA16 Challenge (<https://luna16.grand-challenge.org/results/>). In December of 2017, the ETROCAD system using SOM, ranked 15th out of 18 participants in track 1 of the competition (competition metric: 0.672). The same system, i.e. using the same candidate detector, feature extraction and classification; but using SMOTE or SMOTE-TL as a data balancing method instead of SOM, would achieve a 13th place (competition metric: 0.759) and become the highest ranking system based on handcrafted features (all the systems with higher scores use deep-learning techniques). One should consider the improvement was obtained without adding any additional data or priors to the classification problem.

5.-CONCLUSIONS

In this work, the influence of data balancing techniques on a lung CAD system was evaluated using a web-based framework. Data balancing techniques can be useful to improve the efficiency or boost the performance of classification. The use of data augmentation is always advisable since, although changes in the sensitivity may seem marginal for a big training set, an improvement in the quality of the detection in absolute values can be seen. These methods can have a significant impact on the training process of the SVM, making the generation of classification rules easier. For the application of lung nodule classification, SMOTE and SMOTE-TL were found to perform best, leading to a gain in sensitivity, in particular at lower FP rates. Although, balancing the data can improve the performance of the classification, this is only one factor to take into account in machine learning.

6.-ACKNOWLEDGEMENTS

This work has been supported by the Belgian Development Cooperation through VLIR-UOS (Flemish Interuniversity Council-University Cooperation for Development) in the context of the Institutional University Cooperation Programme with Universidad de Oriente.

Evgenia Papavasileiou is funded by a PhD grant of the Research Foundation Flanders (FWO).

⁵<https://apps.automeris.io/wpd/>

REFERENCIAS

- 1 Tan M, Deklerck R, Jansen B, Bister M, Cornelis J. A novel computer-aided lung nodule detection system for CT images. *MedicalPhysics*.2011;38(10):5631–5645.
- 2 Krawczyk B. Learning from imbalanced data: open challenges and future directions. *Progress in Artificial Intelligence*.2016;5(4):221-232.
- 3 He H, Garcia EA. Learning from imbalanced data. *IEEE Transactions on Knowledge and Data Engineering*. 2009;21(9):1263-1284.
- 4 Krawczyk B. Learning from imbalanced data: open challenges and future directions. *Progress in Artificial Intelligence*. 2016;5(4):221-232.
- 5 Holte RC, Acker L, Porter BW. Concept Learning and the Problem of Small Disjuncts. *Proceedings. International Journal of Conference in Artificial Intelligence*, 1989, 813-818.
- 6 Mease D, Wyner AJ, Buja A. Boosted Classification Tree sand Class Probability/Quantile Estimation. *Journal of Machine Learning Research*. 2007;8:409-439.
- 7 Drummond C, Holte RC. C4.5, Class Imbalance, and Cost Sensitivity: Why Under Sampling Beats Over-Sampling. *Proceedings. International Conference in Machine Learning, Workshop Learning from Imbalanced Data Sets II*. 2003.
- 8 Tomek, I. Two Modifications of CNN. *IEEE Transactions on Systems, Man, and Cybernetics*. 1976;7:679-772
- 9 Zhang, J, Mani, I. KNN Approach to Unbalanced Data Distributions: A Case Study Involving Information Extraction *Proceedings. Workshop on Learning from Imbalanced Datasets*. 2003;Washington DC;USA
- 10 Chawla NV, Bowyer KW, Hall LO, Kegelmeyer, WP. SMOTE: Synthetic Minority Over-sampling Technique. *Journal of Artificial Intelligence Research*. 2002;16(1):321-357
- 11 He H, Bai Y, Garcia EA, Li S. ADASYN: Adaptive Synthetic Sampling Approach for Imbalanced Learning. *International Joint Conference on Neural Networks*. 2008.
- 12 U. of Health., SEER Cancer Statistics Review, 1975-2011.
- 13 de Koning HJ, Meza R, Plevritis SK, Ten Haaf K, Munshi VN, Jeon J et al. Benefits and harms of computed tomography lung cancer screening strategies: a comparative modeling study for the U.S. Preventive Services Task Force. *Annals of Internal Medicine*. 2014;160:311–320
- 14 Black WC, Gareen IF, Soneji SS, Sicks JD, Keeler EB, Aberle DR et al. Cost-effectiveness of CT screening in the National Lung Screening Trial. *N Engl J Med*. 2014;371:1793–1802
- 15 Aberle DR, Henschke CI, McLoud TC, Boiselle PM. Expert opinion: barriers to CT screening for lung cancer. *Journal of Thoracic Imaging*. 2012;27:208
- 16 Prokop M. Lung cancer screening: the radiologist's perspective. *Seminars RespirCrit Care Med*. 2014;35:91–98
- 17 Arindra Adiyoso Setio A, Traverso A, de Bel T, Berens MSN, van den Bogaard C, Cerello P, et al. Validation, comparison, and combination of algorithms for automatic detection of pulmonary nodules in computed tomography images: the luna16 challenge. *Medical Image Analysis (2017)*, 10.1016/j.media.2017.06.015.
- 18 S.G. Armato, et al., The lung image database consortium (LIDC) and image database resource initiative (IDRI): a completed reference database of lung nodules on CT scans, *Med. Phys.* 38 (2011) 915–931.
- 19 Han F, Wang H, Zhang G, Han H, Song B, Li L, et al. Texture Feature Analysis for Computer-Aided Diagnosis on Pulmonary Nodules. *Journal of Digital Imaging*.2015;28:99-115
- 20 Abbas Q. Segmentation of differential structures on computed tomography images for diagnosis lung-related diseases, *Biomedical Signal Processing and Control*.2017;33:325 - 334
- 21 Magdy E, Zayed N, Fakhr, M. Automatic Classification of Normal and Cancer Lung CT Images Using Multiscale AM-FM Features. *Journal of Biomedical Imaging*., 2015:230830:1--230830:7
- 22 Nithila EE, Kumar S. Automatic detection of solitary pulmonary nodules using swarm intelligence optimized neural networks on CT images. *Engineering Science and Technology, an International Journal*.2016
- 23 Madero Orozco H, Vergara Villegas OO, Cruz Sánchez VG, Ochoa Domínguez HJ, Nandayapa Alfaro MJ. Automated system for lung nodules classification based on wavelet feature descriptor and support vector machine. *Biomedical Engineering OnLine*.2015;14:1-20
- 24 Cirujeda P, Muller H, Rubin D, Aguilera TA ,Loo BW, Diehn M, et al. 3D Riesz-wavelet based Covariance descriptors for texture classification of lung nodule tissue in CT. *International Conference of the IEEE Engineering in Medicine and Biology Society*. 2015
- 25 Badura P, Pietka E. Soft computing approach to 3D lung nodule segmentation in CT. *Computers in Biology and Medicine*.2014;53:230-243

- 26 Jacobs C, van Rikxoort EM, Murphy K, Prokop M, Schaefer-Prokop CM, van Ginneken, B. Computer-aided detection of pulmonary nodules: a comparative study using the public LIDC/IDRI database. *European Radiology*.2015;1-9
- 27 Sun T, Zhang R,Wang J, LiX, Guo X. Computer-Aided Diagnosis for Early-Stage Lung Cancer Based on Longitudinal and Balanced Data. *PLoS One*, 2013, 8
- 28 Firmino M, Angelo G, Morais H, Dantas MR, Valentim, R. Computer-aided detection (CADE) and diagnosis (CADx) system for lung cancer with likelihood of malignancy. *Biomedical Engineering OnLine*.2016;15:1-17
- 29 Keshani M, Azimifar Z, Tajeripour F, Boostani R. Lung nodule segmentation and recognition using SVM classifier and active contour modeling: A complete intelligent system. *Computers in biology and medicine*. 2013;43(4):287-300.
- 30 Demir Ö, Çamurcu AY. Computer-aided detection of lung nodules using outer surface features. *Biomedical Materials and Engineering*. IOS Press.2015;26:1213-1222
- 31 Shiju Y, Wei Q,Yubao G and Bin Z. Improving lung cancer prognosis assessment by incorporating synthetic minority oversampling technique and score fusion method *Medical Physics*.2016;43:2694-2703
- 32 Gong J, Liu J-Y, Sun X-W, Zheng B, and Nie, S-D. Computer-aided diagnosis of lung cancer: the effect of training data sets on classification accuracy of lung nodules. *Physics in Medicine & Biology*.2018;63:035036
- 33 Jingchen M, Zien Z, Yacheng R, Junfeng X, Ling F, Qian W and Jun, Z. Computerized detection of lung nodules through radiomics. *Medical Physics*.2017;44:4148-4158
- 34 Demir Ö,Çamurcu AY. Computer-aided detection of lung nodules using outer surface features. *Biomedical Materials and Engineering*. IOS Press.2015;26:1213-1222
- 35 Makaju S, Prasad P, Alsadoon A, Singh A and Elchouemi A. Lung Cancer Detection using CT Scan Images. *Procedia Computer Science*.2018;125:107-114
- 36 Baboo CDS and Iyyapparaj E. Detection Of Classification Methods For Diagnosis Of Pulmonary Nodules In CT Images *IOSR Journal of Computer Engineering*. 2017;19:44-52
- 37 Cao P, Liu X, Zhang J, Zhao D, Huang M and Zaiane O. l2,1 norm regularized multi-kernel based joint nonlinear feature selection and over-sampling for imbalanced data classification. *Neurocomputing*.2017;234:38-57
- 38 Sui Y, Wei Y and Zhao D. Computer-Aided Lung Nodule Recognition by SVM Classifier Based on Combination of Random Undersampling and SMOTE. *Computational and Mathematical Methods in Medicine*.2015;2015:1-13
- 39 Mehre SA, Mukhopadhyay S, Dutta A, Harsha, NC, Dhara AK and Khandelwal N. An automated lung nodule detection system for CT images using synthetic minority oversampling. *Medical Imaging: Computer-Aided Diagnosis, SPIE*. 2016:9785;97850H
- 40 Cai L, Wang X, Wang Y,Guo Y, Yu J, Wang Y. Robust phase-based texture descriptor for classification of breast ultrasound images. *BioMedical Engineering Online*.2015;60:391-402.
- 41 Lemaitre, Guillaume, Fernando Nogueira, and Christos K. Aridas. "Imbalanced-learn: A python toolbox to tackle the curse of imbalanced datasets in machine learning." *Journal of Machine Learning Research*.18.17 (2017): 1-5.

AUTHORS

Alexander Sónora-Mengana, Electronic and Telecommunication Engineer, MSc in Biomedical Engineering, Centro de Biofísica Médica, Santiago de Cuba, Cuba. Main research interests: biomedical signal and image processing, biomedical engineering.

Evgenia Papavasileiou, Electrical and Computer Engineer, MSc in Biomedical Engineering, Vrije Universiteit Brussel, Brussels, Belgium. Main research interests: evolutionary computation, machine learning, biomedical engineering

Juan Carlos-García, Electronic and Telecommunication Engineer, MSc in Biomedical Engineering, Centro de Biofísica Médica, Santiago de Cuba, Cuba. Main research interests: biomedical signal and image processing, biomedical engineering.

Bart Jansen, MSc in Computer Science in Artificial Intelligence, PhD in Sciences. Vrije Universiteit Brussel, Brussels, Belgium. Main research interests: Medical Image Processing, e-Health

Jef Vandemeulebroucke, MSc in Electrical Engineering, PhD in biomedical engineering, Vrije Universiteit Brussel, Brussels, Belgium. Main research interests: medical image analysis, image-guided therapy, computer-aided diagnosis



Los contenidos de la revista se distribuyen bajo una licencia Creative Commons Attribution-NonCommercial 3.0 Unported License